

## DICOM Correction Proposal

STATUS	Letter Ballot
Date of Last Update	2022/06/23
Person Assigned	Bjorn Nolte
Submitter Name	Heinz Blendinger (heinz.blendinger@t-online.de)
Submission Date	2022/01/03

Correction Number	CP-2198
Log Summary:	PS3.16, Correct row reference in Content Item Description of TID 10003
Name of Standard	PS3.16
Rationale for Correction:	When splitting up the former single TID 10003 "Irradiation Event X-Ray Data" by adding sub-templates 10003a to 10003b, a row reference in the Content Item Description was not changed to reflect the appropriate sub-template and related row number.
Correction Wording:	

*Change Row 6 Content Item Description under table "Table TID 10003. Irradiation Event X-Ray Data"*

### Content Item Descriptions

...	
Row 6	The DateTime that the application of X-Rays started for this irradiation event. This shall correspond to the start of the first irradiation in the Irradiation Event, which defines the starting point for the calculation of <b>TID 10003B „Irradiation Event X-Ray Source Data“</b> Row <del>1036</del> "Irradiation Duration".
...	